



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

Mass Flow Technology, Inc.
295 South Post Oak Road, Unit A, Sulphur, LA 70663

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Laboratory and Field Calibration of Flow, Pressure and Temperature Indicating and Recording Devices
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation Body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President/Operations Manager

Initial Accreditation Date:

June 3, 2003

Issue Date:

February 28, 2020

Expiration Date:

December 31, 2021

Accreditation No.:

29291

Certificate No.:

L20-113-2

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver, Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjilabs.com



Certificate of Accreditation: Supplement

Mass Flow Technology, Inc.

295 South Post Oak Road, Unit A, Sulphur, LA 70663
 Contact Name: Eddie Daily Phone: 281-427-7284

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type J ^{FO}	0 °C to 1 200 °C	1 °C	Electrical Simulation of Thermocouple Output Fluke 744 OM 400-6
Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type K ^{FO}	0 °C to 1 370 °C	1 °C	
Temperature Calibration, Indication, and Control Equipment used with Thermocouple Type E ^{FO}	0 °C to 950 °C	1 °C	
Temperature Calibration, Indication, and Control Equipment used with 100 Ohm Pt (385) RTD ^{FO}	0 °C to 800 °C	1 °C	Electrical Simulation of Thermocouple Output Fluke 744 OM 400-5

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Flow Measurement Devices – Water and Process- Mass and/ or Volume – Coriolis, Magnetic, Vortex, Turbine, Rotameters, Paddle Wheels ^F	113 kg/min to 1 130 kg.min (250 lb/min to 2 500 lb/min)	0.1 % of Reading	TSM-4K Transfer Standard Flow Calibration System – Dynamic Comparison Method TSM1: MMI CMF200 OM 400-11, OM 400-12
	113 l/min to 1 130 l/min (30 gal/min to 300 gal/min)		
	227 kg/min to 2 270 kg.min (500 lb/min to 5 000 lb/min)	0.1 % of Reading	TSM-4K Transfer Standard Flow Calibration System – Dynamic Comparison Method TSM1: MMI CMF300 OM 400-11, OM 400-12
	2 27 l/min to 2 270 l/min (60 gal/min to 600 gal/min)		
	454 kg/min to 4 540 kg.min (1 000 lb/min to 10 000 lb/min)	0.1 % of Reading	TSM-4K Transfer Standard Flow Calibration System – Dynamic Comparison Method TSM1: MMI CMF400 OM 400-11, OM 400-12
	454 l/min to 4 540 l/min (120 gal/min to 1 200 gal/min)		



Certificate of Accreditation: Supplement

Mass Flow Technology, Inc.

295 South Post Oak Road, Unit A, Sulphur, LA 70663
Contact Name: Eddie Daily Phone: 281-427-7284

Accreditation is granted to the facility to perform the following calibrations:

Mechanical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Pressure Gauges and Pressure Transmitters ^{FO}	0.00 35 bar to 2.07 bar (0.05 psi to 30 psig)	0.001 bar (0.015 psig)	Fluke 744 w/700P05 Pressure module OM 400-4
	1.7 bar to 690 bar (24 psi to 10 000 psig)	0.55 bar (8 psig)	Fluke-744 w/700P31 Pressure module OM 400-4
	0.69 bar to 207 bar (10 psi to 3 000 psig)	0.17 bar (2.4 psig)	Fluke-744 w/700P29 Pressure module OM 400-4
	0.89 bar to 13.8 bar (-13 psi to 200 psig)	0.01 bar (0.14 psig)	Fluke 744 w/700PD7 Pressure module OM 400-4

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represents the smallest measurement uncertainty attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is typically expressed at a confidence level of 95 % using a coverage factor k (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript FO means that the laboratory performs calibration of the indicated parameter both at its fixed location and onsite at customer locations. Example: Outside Micrometer^{FO} would mean that the laboratory performs this calibration at its fixed location and onsite at customer locations.
4. Measurement uncertainties obtained for calibrations performed at customer sites can be expected to be larger than the measurement uncertainties obtained at the laboratories fixed location for similar calibrations. This is due to the effects of transportation of the standards and equipment and upon environmental conditions at the customer site which are typically not controlled as closely as at the laboratories fixed location.